



Roosin medical Co.,Ltd.

Test Report

Report No.: 2015060101

Name :	Burn gel Dressing	Spec. :	10cm×10cm/40g
Lot :	1505	Sterilization Lot:	20150525
Quantity :	30000 Pieces	Test Quantity :	50 Pieces
Test Standard :	YZB/SU ****-2008	Test objective:	Delivery inspection
Test Date :	May 26th, 2015	Approval Date:	May 26th, 2015

Item	Test Report	Test Result
Appearance	White/Light yellow clear solution	Pass
Specification	Installed Capacity : 36.00g~48.00g Size : (10cm×10cm) ±5%	38.20g~47.69g Long : (9.9cm~10.1cm) Wide:(9.9cm~10.1cm)
PH	pH 5.0-7.5	6.40
Viscosity	5000~50000mPa·s	15433
Sterilization	Gamma Irradiation	Pass

Conclusion : Based on above test results,products under this lot meet the standard

QC : *Liqi Chen*
Date : 2015.06.01

Reviewer : *Yaqin Zhang*
Date : 2015.06.01

QA : *Yangjuan Pan*
Date : 2015.06.01



Roosin Medical Co.,Ltd.

Add.No.8 Yuandong Road,Kouan Town,Gaogang,Taizhou City,Jiangsu Province,P.R.CHINA

TEL:86-523-86908085

FAX:86-523-86908351

Http:www.roosin.com

E-mail:info@roosin.com